

## IRt/c TESTED FOR VACUUM AND MICROWAVE COMPATIBILITY

Tests for compatibility performed by customers contribute greatly to our understanding of how IRt/c's perform in some of the unusual environments where non-contact temperature measurements are required. Exergen has a policy of providing sensors free of charge to any laboratory for test purposes, in return for a copy of test results. This series of tests was performed by a customer who required the performance, and had the facility to test the IRt/c. The test was conducted in a 4 ft. by 4 ft. (1.2 m by 1.2 m) chamber in the following sequence:

### First Test

1. Vacuum exposure (15 minutes at 40 Torr)
2. Microwave Exposure (5 minutes, 3 KW at 2450 MHz)
3. Chamber load (15 pounds H2O)

### Second Test

1. Vacuum exposure (10 minutes)
2. Microwave Exposure (30 seconds, 3 KW at 2450 MHz)
3. Microwave Exposure (10 seconds, 3 KW at 2450 MHz)
4. Chamber load (1 pint H2O)

The IRt/c showed no ill effects and operated flawlessly when checked after the test sequence.

### Exergen Global offices:

USA  
400 Pleasant Street  
Watertown, MA 02472  
Tel: +1 617 923 9900 press 4 for industrial  
Fax: +1 617 923 9911

The Netherlands  
Pastoor Clercxstraat 26  
5465 RH Veghel  
Tel: +31 (0)413 376 599  
Fax: +31 (0)413 379 310

industrial@exergen.com  
www.exergen.com